



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 8:37 am, Jul 08, 2013

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201302	NAME OF AGENCY SAINT JOSEPH POLICE DEPARTMENT	DATE OF INSPECTION 07-04-2013
LOCATION OF INSTRUMENT (STREET AND CITY) 501 FARAON STREET SAINT JOSEPH, MISSOURI 64501		TIME OF INSPECTION 1457

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/04/2013 14:57
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO MARKETING LOT # 12001 EXP. DATE 04-02-2014	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2278 EXP. DATE 12-17-2013	
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 w- .098	TEST 2 a- .098	TEST 3 w .098
--------------------------	--------------------------	-------------------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	4	(.10-.14)	4	(.15-.19)	1	OVER .19	2
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	----------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 1	PRINT FULL NAME OFFICER WAYNE BYROM & OFFICER BRAD KERNS
TYPE 11 PERMIT NUMBER/EXPIRATION DATE 220103 05-09-2014 & 220427 12-27-2014	TELEPHONE NUMBER (816) 271-5359

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901**

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12001

EXPIRATION DATE: April 2, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1205 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 3, 2012.
The expiration date for this lot number is April 2, 2014
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

ICE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201302
07/04/13

REST TIME: 13:00
OBJECT NAME:
KERNS/B/M
OB: 02/02/88 SEX: M
TATE/D.L.: MO/00000
RESTING OFFICER:
KERNS/B/M
FFICER I.D.: 2072
ESTING OFFICER:
BYROM/WAYNE
FFICER I.D.: 7054
ERMIT NUMBER: 220103
XPIRATION DATE: 05/09/14
ISCELLANEOUS DATA:
MAINTENANCE

--- BREATH ANALYSIS ---

ADIOD INTERFERENCE

ck No.
TOR SIGNATURE

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201302
07/04/13
14:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 48c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
ST. JOSEPH POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201302
07/04/13

TESTING OFFICER:
BYROM/WAYNE
OFFICER I.D.: 7054
PERMIT NUMBER: 220103
EXPIRATION DATE: 05/09/14
MISCELLANEOUS DATA:


--- SUPERVISOR MODE ---

BLANK TEST	.000	15:13
INTERNAL STANDARD	VERIFIED	15:13
EXTERNAL STANDARD	.098	15:13
BLANK TEST	.000	15:14
EXTERNAL STANDARD	.098	15:14
BLANK TEST	.000	15:15
EXTERNAL STANDARD	.098	15:15
BLANK TEST	.000	15:16

N = 3

SIM. = .1

AVG. = .098

OPERATOR SIGNATURE 

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



WAYNE BYROM

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER;ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 05/09/2012

Number 220103

Expires 05/09/2014

Director of State Public Health Laboratory

Director, Department of Health

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRAD M KERNS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/27/2012

Number 220427

Expires 12/27/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

ACTING DIRECTOR

Director, Department of Health

Lab. 4 (R7-88)